



防爆合格证

证号: GYJ04327

由 西门子股份公司

(地址: Östliche Rheinbrückenstraße 50, Karlsruhe, Germany)

制造的产品:

名 称 SITRANS P-DSIII FF 压力/差压变送器

型号规格 7MF4 系列

防爆标志 Ex ia IIC T4~T6

产品标准 /

图样编号 A5E00215457B、A5E00216066B

经图样及技术文件的审查和样品检验, 确认上述产品
符合 GB3836.1-2000、GB3836.4-2000 标准,

特颁发此证。有效期自颁发日期起叁年内有效。

备注 1. 产品使用注意事项见本合格证书附件;
2. 关联设备见本合格证书附件。

站 长

国家级仪器仪表防爆安全监督检验站
颁发日期 二〇〇四年九月二十七日



本证书仅对与认可文件和样品一致的产品有效。

地址: 上海市漕宝路103号
邮编: 200233

网址: www.nepsi.org.cn
Email: nepsi@online.sh.cn

电话: 0086 21 64368180
传真: 0086 21 64844580



EXPLOSION PROTECTION

CERTIFICATE OF CONFORMITY

Cert No. GYJ04327

This is to certify that the product

SITRANS P-DSIII FF Pressure/Differential Pressure Transmitter

manufactured by

Siemens Aktiengesellschaft

(Address: Östliche Rheinbrückenstraße 50, Karlsruhe, Germany)

which model is

7MF4 Series

Ex marking

Ex ia II C T4~T6

product standard /

drawing number

A5E00215457B, A5E00216066B

has been inspected and certified by NEPSI, and that it conforms

to GB3836.1-2000 GB3836.4-2000

This Approval shall remain in force until 2007.09.26

Remarks 1. Special conditions for safe use specified in the attachment to this certificate.
2. Associated apparatus specified in the attachment to this certificate.

Director



National Supervision and Inspection Centre for
Explosion Protection and Safety of Instrumentation

Issued Date 2004.09.27

This Certificate is valid for products compatible with the documents and samples approved by NEPSI.

103 Cao Bao Road
Shanghai 200233, China

<http://www.nepsi.org.cn>
Email: nepsi@online.sh.cn

Tel: 0086 21 64368180
Fax: 0086 21 64844580